

检定证书

VERIFICATION CERTIFICATE



证书编号: JL2413795021

第 1 页, 共 3 页

Page 1 of 3 Pages

送检单位

Applicant

深圳大学总医院

计量器具名称

Name of Instrument

医用磁共振成像系统

型号/规格

Type/Specification

prisma

出厂编号

Serial No

166209

资产编号

Asset No

未标明

制造单位

Manufacturer

西门子

检定依据

Verification Regulation

JJG(粤)009-2023 医用磁共振成像系统

检定结论

Conclusion

合格

批准人:

Approved by

谢康

核验员:

Checked by

叶耿

检定员:

Verified by

王恩利

(证书专用章)

Stamp

检定日期:

Operation Date

2024 年 09 月 04 日

Year Month Day

有效期至:

Suggested Recal.Date

2025 年 09 月 03 日

Year Month Day

证书首页背面“重要声明”是证书的组成部分,任何未包含“重要声明”内容的复制均为不完整复制。

法定计量检定机构授权证书号: (粤)法计(2019)01002 号

Authorization Certificate No.

地址: 广东省深圳市南山区龙珠大道 92 号

Address

邮编: 518055

Post Code

客户服务热线: 400 900 8999 - 1

Customer Service Hotline.

电子邮件: kfzx@smq.com.cn

Email

重 要 声 明

Important statement

1. 本院(站)是由深圳市人民政府依据《中华人民共和国计量法》设置并由国家市场监督管理总局、广东省市场监督管理局依法授权的法定计量检定机构。
SMQ is a legal metrological verification organization established by the Shenzhen Municipal People's Government and authorized by the State Administration for Market Regulation and Guangdong Administration for Market Regulation according to the Law on Metrology of the People's Republic of China.
2. 本院(站)进行的检定、校准和检测均可溯源至国际单位制(SI)单位和社会公用计量标准。
All verifications, calibrations and tests made by SMQ are traceable to the International System of Units (SI) and social public measurement standards.
3. 未经本院(站)许可,不得部分复印、摘用或篡改本证书/报告的内容;复印证书/报告未重新加盖本院证书/报告专用章无效。
Copying or excerpting portion of, or altering the content of the certificate/report is not permitted without the written authorization of SMQ. Any copy of certificates/reports without the Dedicated certificate/report Seal is deemed to be invalid.
4. 如果要满足被校准仪器的技术指标,或者技术法规要求,在规定范围内适用,请在建议复校日期前校准(适用于校准报告)。
To ensure that the calibrated object is properly used under given conditions in compliance with technical specifications or regulations, recalibrate before the suggested date (Applicable to calibration report only).
5. 本证书/报告提供的结果仅对本次被检的计量器具有效。
The results provided by the certificate/report are only valid for the measuring object under test this time.
6. 证书/报告无检定员/校准员、核验员、批准人签字,或涂改,或未盖本院证书/报告专用章及骑缝章无效。
Any certificate/report having not been signed by relevant responsible engineer, reviewer or authorized approver, or having been altered without authorization, or without both the Dedicated certificate/report Seal and its across-page seal is deemed to be invalid.
7. 证书/报告更改后,发出的电子版证书/报告、证书/报告的扫描件及传真件将不被追回,委托方有义务将更改后的证书/报告提供给使用原证书/报告的相关方。
SMQ is not responsible for recalling the electronic version of the original certificate/report when any revision is made to them. The applicant assumes the responsibility of providing the revised version to any interested party who uses them.
8. 申领电子证书时,相关内容和效力以电子证书为准;电子证书和纸质证书同时申领时,电子证书仅作为纸质证书的副本,相关内容和效力以同编号纸质证书为准
The relevant content and effectiveness is subject to the electronic version of the original report which was only applied for. When an electronic report and a paper report are applied for at the same time, the electronic report is only a copy of the paper report, and the relevant content and effectiveness is subject to the paper report.
9. 证书中二维码具浏览和下载完整报告功能是应委托方选择所设,该二维码及其复制图能使任何人扫描获取完整的证书电子版,本证书持有人如需限制他人经该二维码获取证书内容,应自行遮盖或消除证书及其复制件所附二维码,我院对委托方选择证书二维码功能所致的信息泄露概不负责(适用于附二维码证书)。
The QR code has the function of browsing and downloading complete report. Setting this function or not is chosen by the customer. The QR code and its copy enable anyone to scan and obtain the complete electronic version of the test report. Thus, if the owner of this report needs to restrict others from obtaining the content of the test report through the QR code, he shall cover or remove the QR code attached to the test report and its copies by himself. SMQ assumes no responsibility for the information leakage caused by the customer's selection of the QR code function of the test report (This clause applies to certificates with QR code attached) ..

获得的国家、省、市专业站 Establishment of Following Institutions

国家数字电子产品质量检验检测中心
National Digital Electronic Product Testing Center (NETC)
国家营养食品质量检验检测中心(广东)
National Nutrition Food Testing Center (Guangdong)
国家体育用品质量检验检测中心(广东)
National Sports Product Testing Center (Guangdong)
国家环保产品质量检验检测中心(广东)
National Environmental Product Testing Center (Guangdong)
国家分布式光伏发电系统质量检验检测中心(广东)
National Distributed Photovoltaic Power Generation System Testing Center (Guangdong)
国家电动汽车电池及充电系统产业计量测试中心
National Metrology Center for Electric Vehicle Battery and Charging System Industry
国家民用无人机产品质量检验检测中心(广东)
National Civil Unmanned Aerial Vehicle Product Testing Center (Guangdong)
国家高新技术计量站
National Hi-tech Metrology Station
国家医疗器械产业计量测试中心
National Metrology and Testing Center for Medical Device Industry
国家计量数据科学中心深圳分中心
National Metrology Data Science Center (Shenzhen)
国家城市能源计量中心(深圳)
National Urban Energy Measurement Center (Shenzhen)
中国轻工业联合会家具质量监督检测深圳站
Accredited Testing Station (Shenzhen) for Furniture by China National Light Industry Council (CNLIC)
广东省质量监督食品检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Food (Shenzhen)
广东省质量监督生态纺织服装产品检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Ecological Textile and Garment Products (Shenzhen)
广东省质量监督皮革制品检验站
Guangdong Quality Supervision and Inspection Institution for Leather Products
广东省质量监督家具检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Furniture Products (Shenzhen)
广东省质量监督学生用品检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Student's Articles (Shenzhen)
广东省质量监督自行车检验站
Guangdong Quality Supervision and Inspection Institution for Bicycle Quality
广东省质量监督钟表检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Horological Products (Shenzhen)
广东省质量监督节能环保产品(安全性能)检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Safety Performance of Environmental Protection and Energy Saving Products (Shenzhen)
广东省质量监督眼镜检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Eyewear Products (Shenzhen)
广东省质量监督电磁兼容检验站
Guangdong Quality Supervision and Inspection Institution for Electromagnetic Compatibility
广东省质量监督综合布线系统检验站
Guangdong Quality Supervision and Inspection Institution for Generic Cabling System
深圳市纤维纺织检验所
Shenzhen Quality Inspection Institution for Fiber and Textile
深圳市建材产品质量监督检验站
Shenzhen Quality Supervision and Inspection Institution for Building Materials
深圳市消防产品燃烧性能检测中心
Shenzhen Testing Center for Burning Behavior of Fire Protection Products

联系方式 Contact Information

联系电话 Contact Tel.: 0755-26941696 26941546(龙珠) 27591789(宝安) 82426246(八卦岭) 28932280(龙岗) 0769-21684525(东莞)
传真电话 Fax No.: 0755-27591716(宝安) 82408176(八卦岭) 28932840(龙岗) 0769-21684527(东莞)
投诉及证书/报告真伪查询电话 Complaint Tel.: 400-900-8999 按5 Email: complaint@smq.com.cn

检定证书

VERIFICATION CERTIFICATE

证书编号: JL2413795021
Certificate No

第 2 页, 共 3 页
Page 2 of 3 Pages

检定用主要计量标准装置信息

Main Standard Devices Used

名称 Equipment Name	测量范围 Measuring Range	不确定度/准确度等级/ 最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	计量标准考核证书号 Certificate No	有效期至 Due Date
医用磁共振成像 系统检定装置	磁场强度: (0.043~3.4) T; 空间分辨力: (1~11)) Lp/cm; 低对比分辨力 (直径): 4.0mm, 6.0mm , 10.0mm; 目标深度: 0.5mm, 0.75mm, 1.0mm , 2.0mm	磁场强度 MPE: $\pm 0.3\%$; 空间分辨力: $U_{rel}=1$ Lp/cm ($k=2$); 低对比分辨力: $U_{rel}=0.01$ mm ($k=2$)	[2013]粤量标鹏法证 字第 109 号	2024-09-27

检定用主要标准器信息

Main Standards of Measurement Used

名称 Equipment Name	测量范围 Measuring Range	不确定度/准确度等级/ 最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	设备编号 Equipment No	证书号/溯源单位 Certificate No/ Traceability to	有效期至 Due Date
磁场强度测量仪	0.043T~3.4T	基本误差 $\leq 5 \times 10^{-5}$ T	SB5947	DCcx2024-02997/ 中国计量院	2025-07-08
磁共振性能模体	空间分辨力:(1~11)LP/cm 空间线性:(20~120)mm 低对比分辨力: ($\varnothing 4.0$ mm 深 0.5 mm 的孔~ $\varnothing 10.0$ mm 深 2.0 mm	低对比分辨力及空间线性: $U=0.01$ mm, $k=2$	SB4694	JL2412446901/深圳检测院	2026-08-12

附加说明

Appended Directions

委托日期: 2024 年 09 月 04 日

Application Date

检定地点: 门诊楼放射科 MR2 室

Operation Location

环境条件: 温度 22.4℃ 相对湿度 50.2%

Operation Environment

检定结果

RESULTS OF VERIFICATION

证书编号: JL2413795021
Certificate No.

第 3 页, 共 3 页
Page 3 of 3 Pages

一、扫描条件

线圈(coil)	头 部 (Head)	采集点数 (Acquisition matrix)	250
脉冲序列 (Pulse sequence)	自旋回波 (SE)	显示矩阵 (Reconstruction matrix)	256×256
重复时间 (TR)	500 ms	视野 (FOV)	250 mm
回波时间 (TE)	30 ms	层厚 (Slice thickness)	5.0 mm
平均次数 (NEX)	2	层间距 (Slice gap)	----

二、检定结果

项目	相位和频率 编码方向	检定结果 [横断面(Trs)]	
信噪比	正方向	SNR	472
	反方向	SNR	465
图像均匀性	正方向	U	96.1%
	反方向	U	97.2%
几何畸变	正方向	L	0.7%
	反方向	L	0.3%
层厚	正方向	d(mm)	5.24
		Δ(mm)	0.24
	反方向	d(mm)	5.19
		Δ(mm)	0.19
空间分辨力	正方向	6.0 Lp/cm	
	反方向	6.0 Lp/cm	
低对比分辨力	正方向	可见孔直径: 4.0 mm, 孔深: 0.5 mm	
	反方向	可见孔直径: 4.0 mm, 孔深: 0.5 mm	

以 下 空 白